

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032624\
 Data File : VX040792.D
 Acq On : 26 Mar 2024 09:59
 Operator : JC/MD
 Sample : VX0326MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0326MBL01

Quant Time: Mar 27 02:40:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	45814	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	82397	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	78228	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	39698	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	38490	49.965	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.920%
35) Dibromofluoromethane	5.379	113	28350	45.982	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.960%
50) Toluene-d8	8.647	98	103108	48.841	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.680%
62) 4-Bromofluorobenzene	11.079	95	40260	48.270	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032624\
Data File : VX040792.D
Acq On : 26 Mar 2024 09:59
Operator : JC/MD
Sample : VX0326MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0326MBL01

Quant Time: Mar 27 02:40:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:44:48 2024
Response via : Initial Calibration

