

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032624\
 Data File : VX040796.D
 Acq On : 26 Mar 2024 12:04
 Operator : JC/MD
 Sample : P1867-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B-148-GW01

Quant Time: Mar 27 02:43:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	45168	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	79719	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	78150	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	38713	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	37342	49.168	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.340%
35) Dibromofluoromethane	5.379	113	27970	46.890	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.780%
50) Toluene-d8	8.646	98	100761	49.333	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.660%
62) 4-Bromofluorobenzene	11.079	95	39000	48.330	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.660%
Target Compounds						
16) Acetone	2.373	43	840	3.123	ug/l	Qvalue # 69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032624\
Data File : VX040796.D
Acq On : 26 Mar 2024 12:04
Operator : JC/MD
Sample : P1867-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B-148-GW01

Quant Time: Mar 27 02:43:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:44:48 2024
Response via : Initial Calibration

