

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032625\
 Data File : VX045461.D
 Acq On : 26 Mar 2025 11:46
 Operator : JC/MD
 Sample : Q1629-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW09-MW01S-20250320

Quant Time: Mar 27 01:34:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	72703	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	139868	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	129727	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	52360	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	58838	50.878	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.760%
35) Dibromofluoromethane	5.373	113	45250	48.382	ug/l	-0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.760%
50) Toluene-d8	8.647	98	159593	47.069	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.140%
62) 4-Bromofluorobenzene	11.079	95	55417	49.323	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032625\
Data File : VX045461.D
Acq On : 26 Mar 2025 11:46
Operator : JC/MD
Sample : Q1629-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW09-MW01S-20250320

Quant Time: Mar 27 01:34:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

