

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032625\
 Data File : VX045464.D
 Acq On : 26 Mar 2025 12:56
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Mar 27 01:37:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	63377	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.756	114	123437	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	112128	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48112	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	58563	58.092	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.180%
35) Dibromofluoromethane	5.379	113	43781	53.043	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.080%
50) Toluene-d8	8.646	98	152557	50.983	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.960%
62) 4-Bromofluorobenzene	11.079	95	54036	54.495	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	109.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032625\
Data File : VX045464.D
Acq On : 26 Mar 2025 12:56
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Mar 27 01:37:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

