

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032625\
 Data File : VX045466.D
 Acq On : 26 Mar 2025 13:43
 Operator : JC/MD
 Sample : Q1629-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT188S1-20250320

Quant Time: Mar 27 01:38:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	68093	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	134396	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120469	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47673	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	57930	53.484	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.960%
35) Dibromofluoromethane	5.379	113	44430	49.440	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.880%
50) Toluene-d8	8.647	98	151229	46.418	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.840%
62) 4-Bromofluorobenzene	11.079	95	51582	47.779	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.560%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	1110	2.209	ug/l	97
44) Trichloroethene	7.123	130	4858	5.316	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032625\
Data File : VX045466.D
Acq On : 26 Mar 2025 13:43
Operator : JC/MD
Sample : Q1629-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT188S1-20250320

Quant Time: Mar 27 01:38:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

