

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032719\
 Data File : VX008503.D
 Acq On : 27 Mar 2019 23:12
 Operator : JC/SP
 Sample : K2137-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SEWER-OUTLET-02

Quant Time: Mar 28 10:20:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X032719W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Mar 27 12:01:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.01	128	34387	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.86	114	198594	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.12	117	173033	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.06	65	88636	30.81	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	102.70%
60) 4-Bromofluorobenzene	11.13	95	86895	28.82	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	96.07%
63) Toluene-d8	8.71	98	251210	30.59	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	101.97%

Target Compounds					Ovalue
15) Acetone	2.44	58	425573	845.758	ug/l 97
25) Chloroform	5.21	83	88495	20.433	ug/l 96
30) 2-Butanone	4.68	43	2820	1.148	ug/l # 93
41) Bromodichloromethane	7.90	83	12928	3.918	ug/l 99
43) Ethyl Acetate	4.83	43	25674	5.806	ug/l # 82
53) Dibromochloromethane	9.58	129	2445	0.997	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032719\
Data File : VX008503.D
Acq On : 27 Mar 2019 23:12
Operator : JC/SP
Sample : K2137-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SEWER-OUTLET-02

Quant Time: Mar 28 10:20:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X032719W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Mar 27 12:01:15 2019
Response via : Initial Calibration

