

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032720\
 Data File : VX015382.D
 Acq On : 27 Mar 2020 15:28
 Operator : JC/SP
 Sample : L2011-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GHD1-WC-L-01

Quant Time: Mar 31 12:19:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1160701	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1889766	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1814886	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	945951	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	766283	46.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.62%	
35) Dibromofluoromethane	5.47	113	589924	48.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.68%	
50) Toluene-d8	8.70	98	2288873	51.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.92%	
62) 4-Bromofluorobenzene	11.12	95	987682	54.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.28%	
Target Compounds						
16) Acetone	2.42	43	24201	3.44	ug/l	95
17) Carbon Disulfide	2.55	76	31473	0.94	ug/l	99
19) Methyl tert-butyl Ether	3.17	73	14789	0.35	ug/l	95
65) Chlorobenzene	10.12	112	20076	0.55	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032720\
Data File : VX015382.D
Acq On : 27 Mar 2020 15:28
Operator : JC/SP
Sample : L2011-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GHD1-WC-L-01

Quant Time: Mar 31 12:19:07 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

