

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032720\
 Data File : VX015386.D
 Acq On : 27 Mar 2020 16:59
 Operator : JC/SP
 Sample : L1997-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE109D2-20200316

Quant Time: Mar 30 04:18:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1123549	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1869334	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1773968	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	947539	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	762165	48.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.20%	
35) Dibromofluoromethane	5.47	113	572970	47.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.90%	
50) Toluene-d8	8.70	98	2238383	50.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.76%	
62) 4-Bromofluorobenzene	11.12	95	973994	54.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.96%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.37	101	12661	1.251	ug/l	92
16) Acetone	2.42	43	16430	2.413	ug/l	97
27) cis-1,2-Dichloroethene	4.58	96	4821	0.354	ug/l	98
44) Trichloroethene	7.19	130	498804	36.251	ug/l	99
64) Tetrachloroethene	9.32	164	8549	0.639	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032720\
Data File : VX015386.D
Acq On : 27 Mar 2020 16:59
Operator : JC/SP
Sample : L1997-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE109D2-20200316

Quant Time: Mar 30 04:18:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

