

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032720\
 Data File : VX015387.D
 Acq On : 27 Mar 2020 17:21
 Operator : JC/SP
 Sample : L1997-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE109D3-20200316

Quant Time: Mar 30 04:24:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1164989	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1873693	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1795934	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	940019	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	761802	46.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.72%	
35) Dibromofluoromethane	5.47	113	587480	49.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.10%	
50) Toluene-d8	8.70	98	2263768	51.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.68%	
62) 4-Bromofluorobenzene	11.12	95	980509	54.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.42%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.37	101	21724	2.070	ug/l	95
12) 1,1-Dichloroethene	2.36	96	4568	0.415	ug/l	87
27) cis-1,2-Dichloroethene	4.56	96	7562	0.535	ug/l	87
32) 1,1,1-Trichloroethane	5.47	97	6793	0.325	ug/l #	52
38) Carbon Tetrachloride	5.76	117	10999	0.588	ug/l	91
44) Trichloroethene	7.19	130	874525	63.408	ug/l	93
49) 1,4-Dioxane	7.73	88	3758	10.628	ug/l #	79
52) Toluene	8.78	92	8465	0.272	ug/l	79
64) Tetrachloroethene	9.32	164	12829	0.947	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032720\
Data File : VX015387.D
Acq On : 27 Mar 2020 17:21
Operator : JC/SP
Sample : L1997-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE109D3-20200316

Quant Time: Mar 30 04:24:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

