

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032720\
 Data File : VX015395.D
 Acq On : 27 Mar 2020 20:23
 Operator : JC/SP
 Sample : L1997-06 100X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D2-20200316

Quant Time: Mar 30 04:47:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1120485	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1812144	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1746918	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	918594	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	745113	47.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.30%	
35) Dibromofluoromethane	5.47	113	572556	49.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.86%	
50) Toluene-d8	8.70	98	2219679	52.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.10%	
62) 4-Bromofluorobenzene	11.12	95	955281	55.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.24%	
Target Compounds						
44) Trichloroethene	7.19	130	723749	54.258	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032720\
Data File : VX015395.D
Acq On : 27 Mar 2020 20:23
Operator : JC/SP
Sample : L1997-06 100X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D2-20200316

Quant Time: Mar 30 04:47:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

