

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032720\
 Data File : VX015399.D
 Acq On : 27 Mar 2020 23:23
 Operator : JC/SP
 Sample : VX0327WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0327WBL02

Quant Time: Mar 30 04:51:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1126987	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1831382	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1709401	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	920347	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	749389	47.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.30%	
35) Dibromofluoromethane	5.47	113	568032	48.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.04%	
50) Toluene-d8	8.70	98	2240353	51.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.96%	
62) 4-Bromofluorobenzene	11.12	95	948660	54.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.32%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032720\
Data File : VX015399.D
Acq On : 27 Mar 2020 23:23
Operator : JC/SP
Sample : VX0327WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0327WBL02

Quant Time: Mar 30 04:51:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

