

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032720\
 Data File : VX015409.D
 Acq On : 28 Mar 2020 03:09
 Operator : JC/SP
 Sample : L2021-04 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-B

Quant Time: Mar 30 05:25:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1108482	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.84	114	1838004	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1731116	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	931616	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.04	65	757720	48.47	ug/l	0.00
Spiked Amount						
			Recovery	=	96.94%	
35) Dibromofluoromethane	5.47	113	580205	49.38	ug/l	0.00
Spiked Amount						
			Recovery	=	98.76%	
50) Toluene-d8	8.70	98	2229670	51.54	ug/l	0.00
Spiked Amount						
			Recovery	=	103.08%	
62) 4-Bromofluorobenzene	11.12	95	958641	54.53	ug/l	0.00
Spiked Amount						
			Recovery	=	109.06%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032720\
Data File : VX015409.D
Acq On : 28 Mar 2020 03:09
Operator : JC/SP
Sample : L2021-04 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-B

Quant Time: Mar 30 05:25:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

