

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032720\
 Data File : VX015415.D
 Acq On : 28 Mar 2020 05:25
 Operator : JC/SP
 Sample : L2033-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ZIPPER-DRAINS

Quant Time: Mar 30 05:37:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1082079	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1790255	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1682437	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	903917	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	732909	48.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.04%	
35) Dibromofluoromethane	5.47	113	569208	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	
50) Toluene-d8	8.70	98	2167673	51.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.90%	
62) 4-Bromofluorobenzene	11.12	95	927140	54.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.30%	
Target Compounds						
16) Acetone	2.42	43	200511	30.581	ug/l	97
43) Isopropyl Acetate	6.42	43	404407	11.154	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032720\
Data File : VX015415.D
Acq On : 28 Mar 2020 05:25
Operator : JC/SP
Sample : L2033-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ZIPPER-DRAINS

Quant Time: Mar 30 05:37:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

