

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032723\
 Data File : VX034738.D
 Acq On : 27 Mar 2023 09:53
 Operator : JC/MD
 Sample : VX0327MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0327MBL01

Quant Time: Mar 27 16:06:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	338022	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	560273	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	492377	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	213688	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	232545	47.289	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.580%
35) Dibromofluoromethane	5.385	113	178326	48.207	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.420%
50) Toluene-d8	8.647	98	670534	48.557	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.120%
62) 4-Bromofluorobenzene	11.079	95	248255	44.507	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032723\
Data File : VX034738.D
Acq On : 27 Mar 2023 09:53
Operator : JC/MD
Sample : VX0327MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0327MBL01

Quant Time: Mar 27 16:06:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

