

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032723\
 Data File : VX034739.D
 Acq On : 27 Mar 2023 10:17
 Operator : JC/MD
 Sample : VX0327WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0327WBL01

Quant Time: Mar 27 16:06:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.550	168	335618	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	565188	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	506161	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	228697	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	234955	48.121	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.240%
35) Dibromofluoromethane	5.385	113	180182	48.285	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.560%
50) Toluene-d8	8.647	98	671451	48.200	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.400%
62) 4-Bromofluorobenzene	11.079	95	259672	46.149	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.300%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032723\
 Data File : VX034739.D
 Acq On : 27 Mar 2023 10:17
 Operator : JC/MD
 Sample : VX0327WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0327WBL01

Quant Time: Mar 27 16:06:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

