

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032723\
 Data File : VX034742.D
 Acq On : 27 Mar 2023 11:34
 Operator : JC/MD
 Sample : 02045-06
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC1-20230322

Quant Time: Mar 27 16:07:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	331221	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	553744	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	495461	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	222795	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	229669	47.663	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.320%
35) Dibromofluoromethane	5.385	113	165245	45.197	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.400%
50) Toluene-d8	8.647	98	662820	48.564	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.120%
62) 4-Bromofluorobenzene	11.085	95	248700	45.112	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032723\
Data File : VX034742.D
Acq On : 27 Mar 2023 11:34
Operator : JC/MD
Sample : 02045-06
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC1-20230322

Quant Time: Mar 27 16:07:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

