

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032723\
 Data File : VX034746.D
 Acq On : 27 Mar 2023 13:30
 Operator : JC/MD
 Sample : 02104-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-2023-03-25

Quant Time: Mar 28 03:18:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	336476	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	563160	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	507663	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	228846	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	237214	48.460	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.920%
35) Dibromofluoromethane	5.385	113	176630	47.504	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.000%
50) Toluene-d8	8.647	98	675476	48.664	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.320%
62) 4-Bromofluorobenzene	11.079	95	259982	46.370	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032723\
Data File : VX034746.D
Acq On : 27 Mar 2023 13:30
Operator : JC/MD
Sample : 02104-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-2023-03-25

Quant Time: Mar 28 03:18:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

