

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032723\
 Data File : VX034748.D
 Acq On : 27 Mar 2023 14:39
 Operator : JC/MD
 Sample : 02100-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DS-01-032323

Quant Time: Mar 28 03:19:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	311817	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.763	114	524593	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	465700	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	205756	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	220693	48.650	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.300%
35) Dibromofluoromethane	5.385	113	161848	46.728	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.460%
50) Toluene-d8	8.646	98	620692	48.004	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.000%
62) 4-Bromofluorobenzene	11.079	95	236347	45.254	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032723\
Data File : VX034748.D
Acq On : 27 Mar 2023 14:39
Operator : JC/MD
Sample : 02100-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DS-01-032323

Quant Time: Mar 28 03:19:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

