

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032723\
 Data File : VX034755.D
 Acq On : 27 Mar 2023 17:22
 Operator : JC/MD
 Sample : 02073-04 1000X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRE-1660

Quant Time: Mar 28 03:21:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	239932	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.763	114	408670	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	355314	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	154286	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	159935	45.820	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.640%
35) Dibromofluoromethane	5.385	113	123979	45.948	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.900%
50) Toluene-d8	8.647	98	480541	47.707	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.420%
62) 4-Bromofluorobenzene	11.079	95	164511	40.434	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	80.860%#
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	299728	182.293	ug/l	99
52) Toluene	8.720	92	27554	3.671	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032723\
Data File : VX034755.D
Acq On : 27 Mar 2023 17:22
Operator : JC/MD
Sample : 02073-04 1000X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRE-1660

Quant Time: Mar 28 03:21:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

