

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032723\
 Data File : VX034759.D
 Acq On : 27 Mar 2023 18:55
 Operator : JC/MD
 Sample : 02073-01 400X
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRE-1715

Quant Time: Mar 28 03:23:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	285946	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.763	114	474720	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	419327	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	196830	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	196561	47.251	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	94.500%		
35) Dibromofluoromethane	5.385	113	144392	46.068	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.140%		
50) Toluene-d8	8.647	98	570581	48.765	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.520%		
62) 4-Bromofluorobenzene	11.079	95	210719	44.586	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	89.180%		
Target Compounds						Qvalue
16) Acetone	2.385	43	6222	3.175	ug/l	98
89) n-Butylbenzene	12.335	91	14614	1.244	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032723\
Data File : VX034759.D
Acq On : 27 Mar 2023 18:55
Operator : JC/MD
Sample : 02073-01 400X
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRE-1715

Quant Time: Mar 28 03:23:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

