



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032723\
Data File : VX034736.D
Acq On : 27 Mar 2023 08:56
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Title : SW846 8260
Last Update : Mon Mar 13 10:29:32 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.3	7224	PASS
75	95	30	60	54.5	20379	PASS
95	95	100	100	100.0	37403	PASS
96	95	5	9	6.7	2494	PASS
173	174	0.00	2	0.4	104	PASS
174	95	50	100	70.1	26224	PASS
175	174	5	9	7.3	1903	PASS
176	174	95	101	96.9	25416	PASS
177	176	5	9	7.0	1782	PASS