

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032724\
 Data File : VX040817.D
 Acq On : 27 Mar 2024 16:55
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Mar 28 03:57:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	45676	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	77264	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	68490	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	35997	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	38992	50.769	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.540%	
35) Dibromofluoromethane	5.385	113	27577	47.700	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.400%	
50) Toluene-d8	8.647	98	96201	48.597	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.200%	
62) 4-Bromofluorobenzene	11.079	95	35394	45.255	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 90.500%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.167	85	27096	45.841	ug/l	97
3) Chloromethane	1.301	50	26364	43.608	ug/l	97
4) Vinyl Chloride	1.374	62	28020	48.946	ug/l	99
5) Bromomethane	1.599	94	17695	45.853	ug/l	99
6) Chloroethane	1.685	64	17821	44.289	ug/l	98
7) Trichlorofluoromethane	1.886	101	44762	44.688	ug/l	96
8) Diethyl Ether	2.130	74	14409	44.753	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.331	101	25781	44.759	ug/l	98
10) Methyl Iodide	2.453	142	33622	44.087	ug/l	99
11) Tert butyl alcohol	2.953	59	32829	266.932	ug/l #	90
12) 1,1-Dichloroethene	2.319	96	23720	43.966	ug/l	98
13) Acrolein	2.233	56	18471	253.786	ug/l	95
14) Allyl chloride	2.660	41	43865	45.366	ug/l	90
15) Acrylonitrile	3.056	53	74497	245.371	ug/l	97
16) Acetone	2.374	43	72239	265.549	ug/l	99
17) Carbon Disulfide	2.514	76	58618	37.825	ug/l	98
18) Methyl Acetate	2.703	43	33798	46.357	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	87840	47.173	ug/l	97
20) Methylene Chloride	2.788	84	26964	45.345	ug/l	99
21) trans-1,2-Dichloroethene	3.093	96	24551	43.106	ug/l	94
22) Diisopropyl ether	3.757	45	89016	47.179	ug/l	98
23) Vinyl Acetate	3.721	43	430806	241.556	ug/l	99
24) 1,1-Dichloroethane	3.605	63	49621	45.788	ug/l	99
25) 2-Butanone	4.550	43	111528	261.748	ug/l	99
26) 2,2-Dichloropropane	4.471	77	42430	44.874	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	30559	46.319	ug/l	100
28) Bromochloromethane	4.891	49	23383	47.177	ug/l	98
29) Tetrahydrofuran	4.995	42	73720	260.113	ug/l	99
30) Chloroform	5.087	83	55289	47.654	ug/l	99
31) Cyclohexane	5.465	56	39550	43.341	ug/l	98
32) 1,1,1-Trichloroethane	5.379	97	48284	48.417	ug/l	99
36) 1,1-Dichloropropene	5.690	75	36348	42.700	ug/l	99
37) Ethyl Acetate	4.709	43	45446	47.016	ug/l	97
38) Carbon Tetrachloride	5.672	117	40675	42.453	ug/l	97
39) Methylcyclohexane	7.379	83	42344	44.901	ug/l	96
40) Benzene	6.031	78	103456	44.014	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032724\
 Data File : VX040817.D
 Acq On : 27 Mar 2024 16:55
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Mar 28 03:57:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	24296	48.607	ug/l	96
42) 1,2-Dichloroethane	6.080	62	45027	47.339	ug/l	98
43) Isopropyl Acetate	6.336	43	72644	47.859	ug/l	97
44) Trichloroethene	7.123	130	26193	46.491	ug/l	94
45) 1,2-Dichloropropane	7.428	63	26138	47.396	ug/l	97
46) Dibromomethane	7.574	93	21292	48.748	ug/l	99
47) Bromodichloromethane	7.818	83	44207	49.533	ug/l	97
48) Methyl methacrylate	7.690	41	36174	52.142	ug/l	94
49) 1,4-Dioxane	7.653	88	13269	1133.283	ug/l #	97
51) 4-Methyl-2-Pentanone	8.568	43	241828	256.903	ug/l	97
52) Toluene	8.714	92	65266	46.557	ug/l	94
53) t-1,3-Dichloropropene	8.976	75	43548	47.169	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	43988	46.069	ug/l	99
55) 1,1,2-Trichloroethane	9.147	97	26926	47.166	ug/l	93
56) Ethyl methacrylate	9.116	69	47340	49.611	ug/l #	92
57) 1,3-Dichloropropane	9.305	76	45901	47.118	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	114036	227.708	ug/l	96
59) 2-Hexanone	9.427	43	189367	259.692	ug/l	95
60) Dibromochloromethane	9.519	129	32304	48.154	ug/l	100
61) 1,2-Dibromoethane	9.610	107	28279	46.910	ug/l	99
64) Tetrachloroethene	9.269	164	22337	43.784	ug/l	96
65) Chlorobenzene	10.080	112	71174	45.818	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	24875	47.057	ug/l	99
67) Ethyl Benzene	10.189	91	128195	46.080	ug/l	99
68) m/p-Xylenes	10.299	106	93569	91.865	ug/l	93
69) o-Xylene	10.640	106	44671	45.653	ug/l	89
70) Styrene	10.653	104	80367	49.064	ug/l	97
71) Bromoform	10.799	173	20754	48.070	ug/l #	99
73) Isopropylbenzene	10.964	105	120025	43.944	ug/l	99
74) N-amyl acetate	10.842	43	62954	46.813	ug/l	96
75) 1,1,2,2-Tetrachloroethane	11.213	83	39661	42.893	ug/l	98
76) 1,2,3-Trichloropropane	11.238	75	46465	60.017	ug/l	90
77) Bromobenzene	11.195	156	28164	42.600	ug/l	100
78) n-propylbenzene	11.305	91	145735	43.274	ug/l	97
79) 2-Chlorotoluene	11.366	91	83956	43.204	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	101035	44.125	ug/l	97
81) trans-1,4-Dichloro-2-b...	11.018	75	13564	44.322	ug/l	96
82) 4-Chlorotoluene	11.451	91	101285	41.556	ug/l	99
83) tert-Butylbenzene	11.713	119	99475	43.315	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	101510	44.048	ug/l	98
85) sec-Butylbenzene	11.890	105	128178	43.891	ug/l	99
86) p-Isopropyltoluene	12.006	119	108298	44.721	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	55750	44.158	ug/l	99
88) 1,4-Dichlorobenzene	12.043	146	56768	42.339	ug/l	100
89) n-Butylbenzene	12.329	91	107378	43.971	ug/l	96
90) Hexachloroethane	12.536	117	19102	45.300	ug/l	93
91) 1,2-Dichlorobenzene	12.335	146	54413	43.868	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	9658	41.832	ug/l	94
93) 1,2,4-Trichlorobenzene	13.585	180	37742	42.813	ug/l	98
94) Hexachlorobutadiene	13.725	225	14652	42.167	ug/l	99
95) Naphthalene	13.774	128	127359	45.665	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	37622	43.438	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032724\
Data File : VX040817.D
Acq On : 27 Mar 2024 16:55
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Quant Time: Mar 28 03:57:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:44:48 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:168 Resp: 45676
Ion Ratio Lower Upper
168 100
99 69.4 56.6 85.0



#2
Dichlorodifluoromethane
Concen: 45.841 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.001 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 85 Resp: 27096
Ion Ratio Lower Upper
85 100
87 31.8 16.7 50.1









#3
Chloromethane
Concen: 43.608 ug/l
RT: 1.301 min Scan# 35
Delta R.T. 0.007 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 50 Resp: 26364
Ion Ratio Lower Upper
50 100
52 33.6 25.5 38.3



#4
Vinyl Chloride
Concen: 48.946 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 62 Resp: 28020
Ion Ratio Lower Upper
62 100
64 31.1 25.4 38.0









#5
Bromomethane
Concen: 45.853 ug/l
RT: 1.599 min Scan# 84
Delta R.T. -0.006 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 94 Resp: 17695
Ion Ratio Lower Upper
94 100
96 92.0 74.6 111.8



#6
Chloroethane
Concen: 44.289 ug/l
RT: 1.685 min Scan# 98
Delta R.T. -0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 64 Resp: 17821
Ion Ratio Lower Upper
64 100
66 32.3 25.1 37.7









#7
Trichlorofluoromethane
Concen: 44.688 ug/l
RT: 1.886 min Scan# 131
Delta R.T. -0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 44762
Ion Ratio Lower Upper
101 100
103 67.8 51.4 77.2



#8
Diethyl Ether
Concen: 44.753 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 74 Resp: 14409
Ion Ratio Lower Upper
74 100
45 100.7 53.3 159.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 44.759 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 25781
Ion Ratio Lower Upper
101 100
85 48.6 37.0 55.6
151 73.8 58.6 87.8



#10
Methyl Iodide
Concen: 44.087 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion:142 Resp: 33622
Ion Ratio Lower Upper
142 100
127 51.3 41.6 62.4
141 13.7 11.6 17.4









#11
Tert butyl alcohol
Concen: 266.932 ug/l
RT: 2.953 min Scan# 305
Delta R.T. 0.006 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 59 Resp: 32829
Ion Ratio Lower Upper
59 100
57 10.9 6.0 9.0#



#12
1,1-Dichloroethene
Concen: 43.966 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 96 Resp: 23720
Ion Ratio Lower Upper
96 100
61 178.5 144.6 216.8
98 65.3 50.4 75.6





#13

Acrolein

Concen: 253.786 ug/l

RT: 2.233 min Scan# 188

Delta R.T. 0.000 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 56 Resp: 18471

Ion Ratio Lower Upper

56 100

55 75.2 57.1 85.7



#14

Allyl chloride

Concen: 45.366 ug/l

RT: 2.660 min Scan# 258

Delta R.T. -0.006 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55




Tgt Ion: 41 Resp: 43865


Ion Ratio Lower Upper

41 100

39 80.0 55.8 83.6

76 32.3 24.3 36.5









#15
Acrylonitrile
Concen: 245.371 ug/l
RT: 3.056 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 53 Resp: 74497
Ion Ratio Lower Upper
53 100
52 85.2 66.0 99.0
51 39.2 29.8 44.6



#16
Acetone
Concen: 265.549 ug/l
RT: 2.374 min Scan# 211
Delta R.T. -0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 43 Resp: 72239
Ion Ratio Lower Upper
43 100
58 28.1 21.9 32.9









#17
Carbon Disulfide
Concen: 37.825 ug/l
RT: 2.514 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 76 Resp: 58618
Ion Ratio Lower Upper
76 100
78 9.6 7.0 10.6



#18
Methyl Acetate
Concen: 46.357 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.006 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 43 Resp: 33798
Ion Ratio Lower Upper
43 100
74 21.8 17.8 26.8









#19
Methyl tert-butyl Ether
Concen: 47.173 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 73 Resp: 87840
Ion Ratio Lower Upper
73 100
57 20.2 17.3 25.9



#20
Methylene Chloride
Concen: 45.345 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 84 Resp: 26964
Ion Ratio Lower Upper
84 100
49 128.4 103.6 155.4
51 39.0 33.8 50.6
86 61.8 49.4 74.0









#21
trans-1,2-Dichloroethene
Concen: 43.106 ug/l
RT: 3.093 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 96 Resp: 24551
Ion Ratio Lower Upper
96 100
61 140.6 119.2 178.8
98 63.1 48.7 73.1



#22
Diisopropyl ether
Concen: 47.179 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.006 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 45 Resp: 89016
Ion Ratio Lower Upper
45 100
43 84.9 69.7 104.5
87 25.6 20.8 31.2
59 11.6 9.0 13.4









#23
 Vinyl Acetate
 Concen: 241.556 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. 0.006 min
 Lab File: VX040817.D
 Acq: 27 Mar 2024 16:55




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 43 Resp: 430806
 Ion Ratio Lower Upper
 43 100
 86 9.5 7.8 11.8



#24
 1,1-Dichloroethane
 Concen: 45.788 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: VX040817.D
 Acq: 27 Mar 2024 16:55

Tgt Ion: 63 Resp: 49621
 Ion Ratio Lower Upper
 63 100
 98 6.4 3.3 9.8
 100 3.8 2.1 6.5









#25
2-Butanone
Concen: 261.748 ug/l
RT: 4.550 min Scan# 568
Delta R.T. 0.006 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 111528
Ion Ratio Lower Upper
43 100
72 24.0 19.4 29.2



#26
2,2-Dichloropropane
Concen: 44.874 ug/l
RT: 4.471 min Scan# 555
Delta R.T. -0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 77 Resp: 42430
Ion Ratio Lower Upper
77 100
97 20.9 10.8 32.3





#27

cis-1,2-Dichloroethene

Concen: 46.319 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.000 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 96 Resp: 30559

Ion Ratio Lower Upper

96 100

61 155.9 0.0 312.2

98 63.7 0.0 129.0



#28

Bromochloromethane

Concen: 47.177 ug/l

RT: 4.891 min Scan# 624

Delta R.T. 0.000 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55




Tgt Ion: 49 Resp: 23383


Ion Ratio Lower Upper

49 100

129 1.4 0.0 3.8

130 73.8 57.5 86.3





#29

Tetrahydrofuran

Concen: 260.113 ug/l

RT: 4.995 min Scan# 64

Delta R.T. 0.006 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 42 Resp: 73720

Ion Ratio Lower Upper

42 100

72 42.1 34.2 51.4

71 40.0 32.2 48.2



#30

Chloroform

Concen: 47.654 ug/l

RT: 5.087 min Scan# 656

Delta R.T. -0.006 min

Lab File: VX040817.D




Acq: 27 Mar 2024 16:55


Tgt Ion: 83 Resp: 55289

Ion Ratio Lower Upper

83 100

85 63.4 50.3 75.5









#31
Cyclohexane
Concen: 43.341 ug/l
RT: 5.465 min Scan# 718
Delta R.T. 0.001 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:	56	Resp:	39550
Ion	Ratio	Lower	Upper
56	100		
69	30.9	23.4	35.2
84	84.3	65.9	98.9



#32
1,1,1-Trichloroethane
Concen: 48.417 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.006 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion:	97	Resp:	48284
Ion	Ratio	Lower	Upper
97	100		
99	64.7	51.1	76.7
61	46.5	36.4	54.6









#33
1,2-Dichloroethane-d4
Concen: 50.769 ug/l
RT: 5.952 min Scan# 798
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 65 Resp: 38992
Ion Ratio Lower Upper
65 100
67 48.3 0.0 100.2



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. -0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 114 Resp: 77264
Ion Ratio Lower Upper
114 100
63 24.3 0.0 45.0
88 16.5 0.0 36.6





#35

Dibromofluoromethane

Concen: 47.700 ug/l

RT: 5.385 min Scan# 704

Delta R.T. 0.006 min

Lab File: VX040817.D


Acq: 27 Mar 2024 16:55

Instrument :




MSVOA_X

ClientSampleId :

VSTDCCC050EC



Tgt Ion:	113	Resp:	27577
Ion	Ratio	Lower	Upper
113	100		
111	101.1	82.9	124.3
192	17.0	13.3	19.9



#36

1,1-Dichloropropene

Concen: 42.700 ug/l




RT: 5.690 min Scan# 755


Delta R.T. 0.000 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55

Tgt Ion:	75	Resp:	36348
Ion	Ratio	Lower	Upper
75	100		
110	32.9	17.0	50.9
77	30.5	24.7	37.1





#37

Ethyl Acetate

Concen: 47.016 ug/l

RT: 4.709 min Scan# 594

Delta R.T. 0.006 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 43 Resp: 45446

Ion Ratio Lower Upper

43 100

61 12.4 10.6 16.0

70 9.2 8.3 12.5



#38

Carbon Tetrachloride

Concen: 42.453 ug/l

RT: 5.672 min Scan# 752

Delta R.T. 0.000 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55




Tgt Ion:117 Resp: 40675


Ion Ratio Lower Upper

117 100

119 94.4 78.2 117.2

121 30.5 24.4 36.6









#39
Methylcyclohexane
Concen: 44.901 ug/l
RT: 7.379 min Scan# 1032
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 42344
Ion Ratio Lower Upper
83 100
55 78.5 64.6 97.0
98 50.0 36.7 55.1



#40
Benzene
Concen: 44.014 ug/l
RT: 6.031 min Scan# 811
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 78 Resp: 103456
Ion Ratio Lower Upper
78 100
77 21.5 19.3 28.9









#41
Methacrylonitrile
Concen: 48.607 ug/l
RT: 4.916 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:	41	Resp:	24296
Ion	Ratio	Lower	Upper
41	100		
39	59.5	45.2	67.8
67	66.5	52.6	78.8
52	33.0	22.7	34.1



#42
1,2-Dichloroethane
Concen: 47.339 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion:	62	Resp:	45027
Ion	Ratio	Lower	Upper
62	100		
98	7.4	0.0	16.0









#43
Isopropyl Acetate
Concen: 47.859 ug/l
RT: 6.336 min Scan# 861
Delta R.T. 0.006 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 72644
Ion Ratio Lower Upper
43 100
61 17.0 15.2 22.8
87 12.0 9.8 14.6



#44
Trichloroethene
Concen: 46.491 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion:130 Resp: 26193
Ion Ratio Lower Upper
130 100
95 108.6 0.0 204.4









#45
1,2-Dichloropropane
Concen: 47.396 ug/l
RT: 7.428 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 63 Resp: 26138
Ion Ratio Lower Upper
63 100
65 32.2 24.3 36.5



#46
Dibromomethane
Concen: 48.748 ug/l
RT: 7.574 min Scan# 1064
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 93 Resp: 21292
Ion Ratio Lower Upper
93 100
95 81.8 66.6 100.0
174 86.1 69.4 104.0









#47
Bromodichloromethane
Concen: 49.533 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 44207
Ion Ratio Lower Upper
83 100
85 64.2 49.8 74.6
127 8.0 7.1 10.7



#48
Methyl methacrylate
Concen: 52.142 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.006 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 41 Resp: 36174
Ion Ratio Lower Upper
41 100
69 76.0 59.2 88.8
39 68.0 47.7 71.5









#49
1,4-Dioxane
Concen: 1133.283 ug/l
RT: 7.653 min Scan# 1077
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 88 Resp: 13269
Ion Ratio Lower Upper
88 100
43 39.8 0.0 0.0#
58 72.6 60.1 90.1



#50
Toluene-d8
Concen: 48.597 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 98 Resp: 96201
Ion Ratio Lower Upper
98 100
100 64.7 53.1 79.7









#51
4-Methyl-2-Pentanone
Concen: 256.903 ug/l
RT: 8.568 min Scan# 1227
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 241828
Ion Ratio Lower Upper
43 100
58 36.0 30.4 45.6



#52
Toluene
Concen: 46.557 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 92 Resp: 65266
Ion Ratio Lower Upper
92 100
91 175.2 134.2 201.2









#53
t-1,3-Dichloropropene
Concen: 47.169 ug/l
RT: 8.976 min Scan# 1194
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 75 Resp: 43548
Ion Ratio Lower Upper
75 100
77 31.9 25.7 38.5



#54
cis-1,3-Dichloropropene
Concen: 46.069 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. 0.006 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 75 Resp: 43988
Ion Ratio Lower Upper
75 100
77 31.0 25.3 37.9
39 58.0 46.7 70.1





#55

1,1,2-Trichloroethane

Concen: 47.166 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. 0.000 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 97 Resp: 26926





Ion Ratio Lower Upper

97 100

83 87.9 65.8 98.8

85 58.2 41.6 62.4

99 63.6 47.8 71.8



#56

Ethyl methacrylate

Concen: 49.611 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.006 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55




Tgt Ion: 69 Resp: 47340


Ion Ratio Lower Upper

69 100

41 74.3 56.8 85.2

39 49.0 32.0 48.0#





#57

1,3-Dichloropropane

Concen: 47.118 ug/l

RT: 9.305 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 76 Resp: 45901

Ion Ratio Lower Upper

76 100

78 32.0 26.1 39.1



#58

2-Chloroethyl Vinyl ether

Concen: 227.708 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX040817.D




Acq: 27 Mar 2024 16:55


Tgt Ion: 63 Resp: 114036

Ion Ratio Lower Upper

63 100

106 26.2 22.5 33.7









#59
2-Hexanone
Concen: 259.692 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 189367
Ion Ratio Lower Upper
43 100
58 48.5 26.2 78.5



#60
Dibromochloromethane
Concen: 48.154 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 129 Resp: 32304
Ion Ratio Lower Upper
129 100
127 78.0 38.8 116.4









#61
1,2-Dibromoethane
Concen: 46.910 ug/l
RT: 9.610 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:107 Resp: 28279
Ion Ratio Lower Upper
107 100
109 94.7 75.2 112.8



#62
4-Bromofluorobenzene
Concen: 45.255 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 95 Resp: 35394
Ion Ratio Lower Upper
95 100
174 76.3 0.0 135.6
176 72.3 0.0 131.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1470
Delta R.T. 0.006 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:117 Resp: 68490
Ion Ratio Lower Upper
117 100
82 56.4 46.2 69.4
119 32.2 25.5 38.3



#64
Tetrachloroethene
Concen: 43.784 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion:164 Resp: 22337
Ion Ratio Lower Upper
164 100
166 130.1 102.7 154.1
129 105.4 80.4 120.6
131 102.8 76.5 114.7





#65

Chlorobenzene

Concen: 45.818 ug/l

RT: 10.080 min Scan# 1474

Delta R.T. 0.006 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion:112 Resp: 71174

Ion Ratio Lower Upper

112 100

114 32.8 25.7 38.5



#66

1,1,1,2-Tetrachloroethane

Concen: 47.057 ug/l

RT: 10.159 min Scan# 1488

Delta R.T. 0.000 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55




Tgt Ion:131 Resp: 24875


Ion Ratio Lower Upper

131 100

133 94.8 47.1 141.4

119 65.7 32.2 96.6









#67
Ethyl Benzene
Concen: 46.080 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 128195
Ion Ratio Lower Upper
91 100
106 30.4 24.9 37.3



#68
m/p-Xylenes
Concen: 91.865 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 106 Resp: 93569
Ion Ratio Lower Upper
106 100
91 217.4 165.5 248.3









#69
o-Xylene
Concen: 45.653 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:106 Resp: 44671
Ion Ratio Lower Upper
106 100
91 235.3 108.9 326.8



#70
Styrene
Concen: 49.064 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion:104 Resp: 80367
Ion Ratio Lower Upper
104 100
78 58.0 43.6 65.4
103 55.2 43.6 65.4









#71
Bromoform
Concen: 48.070 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:173	Resp:	20754
Ion Ratio	Lower	Upper
173	100	
175	47.8	24.1 72.4
254	0.0	0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion:152	Resp:	35997
Ion Ratio	Lower	Upper
152	100	
115	85.7	43.3 129.9
150	171.4	0.0 345.0





#73

Isopropylbenzene

Concen: 43.944 ug/l

RT: 10.964 min Scan# 1619

Delta R.T. 0.006 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion:105 Resp: 120025

Ion Ratio Lower Upper

105 100

120 26.0 12.9 38.6



#74

N-ethyl acetate

Concen: 46.813 ug/l

RT: 10.842 min Scan# 1600

Delta R.T. 0.000 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55

Tgt Ion: 43 Resp: 62954




Ion Ratio Lower Upper


43 100

70 37.5 32.9 49.3

55 22.1 19.3 28.9

61 22.2 18.8 28.2





#75

1,1,2,2-Tetrachloroethane

Concen: 42.893 ug/l

RT: 11.213 min Scan# 1

Delta R.T. 0.006 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 83 Resp: 39661

Ion Ratio Lower Upper

83 100

131 9.3 5.3 15.8

85 65.1 32.0 96.2



#76

1,2,3-Trichloropropane

Concen: 60.017 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. 0.000 min

Lab File: VX040817.D




Acq: 27 Mar 2024 16:55


Tgt Ion: 75 Resp: 46465

Ion Ratio Lower Upper

75 100

77 32.5 19.1 57.5





#77

Bromobenzene

Concen: 42.600 ug/l

RT: 11.195 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:156 Resp: 28164

Ion Ratio Lower Upper

156 100

77 176.7 88.4 265.3

158 96.6 48.0 144.0



#78

n-propylbenzene

Concen: 43.274 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. 0.000 min

Lab File: VX040817.D




Acq: 27 Mar 2024 16:55


Tgt Ion: 91 Resp: 145735

Ion Ratio Lower Upper

91 100

120 21.6 11.5 34.4





#79

2-Chlorotoluene

Concen: 43.204 ug/l

RT: 11.366 min Scan# 1

Delta R.T. 0.006 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 91 Resp: 83956

Ion Ratio Lower Upper

91 100

126 32.7 16.3 48.8



#80

1,3,5-Trimethylbenzene

Concen: 44.125 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX040817.D




Acq: 27 Mar 2024 16:55


Tgt Ion:105 Resp: 101035

Ion Ratio Lower Upper

105 100

120 46.1 24.1 72.3









#81
trans-1,4-Dichloro-2-butene
Concen: 44.322 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 75 Resp: 13564
Ion Ratio Lower Upper
75 100
53 117.2 90.5 135.7
89 49.3 36.9 55.3



#82
4-Chlorotoluene
Concen: 41.556 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 91 Resp: 101285
Ion Ratio Lower Upper
91 100
126 28.0 14.2 42.6





#83

tert-Butylbenzene

Concen: 43.315 ug/l

RT: 11.713 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:119 Resp: 99475

Ion Ratio Lower Upper

119 100

91 64.2 31.2 93.6

134 22.3 11.7 35.0



#84

1,2,4-Trimethylbenzene

Concen: 44.048 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX040817.D




Acq: 27 Mar 2024 16:55


Tgt Ion:105 Resp: 101510

Ion Ratio Lower Upper

105 100

120 43.9 21.2 63.6









#85
sec-Butylbenzene
Concen: 43.891 ug/l
RT: 11.890 min Scan# 11772
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:105 Resp: 128178
Ion Ratio Lower Upper
105 100
134 19.1 9.8 29.4



#86
p-Isopropyltoluene
Concen: 44.721 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion:119 Resp: 108298
Ion Ratio Lower Upper
119 100
134 25.9 12.8 38.3
91 27.6 13.1 39.3





#87

1,3-Dichlorobenzene

Concen: 44.158 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:146 Resp: 55750

Ion Ratio Lower Upper

146 100

111 43.1 22.1 66.1

148 64.4 31.8 95.3



#88

1,4-Dichlorobenzene

Concen: 42.339 ug/l

RT: 12.043 min Scan# 1797

Delta R.T. 0.006 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55




Tgt Ion:146 Resp: 56768


Ion Ratio Lower Upper

146 100

111 42.7 21.3 63.7

148 63.8 31.8 95.4









#89
n-Butylbenzene
Concen: 43.971 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 107378
Ion Ratio Lower Upper
91 100
92 51.2 27.4 82.0
134 23.9 12.6 37.8



#90
Hexachloroethane
Concen: 45.300 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 117 Resp: 19102
Ion Ratio Lower Upper
117 100
201 71.1 32.9 98.6









#91
1,2-Dichlorobenzene
Concen: 43.868 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:146 Resp: 54413
Ion Ratio Lower Upper
146 100
111 44.3 22.4 67.0
148 62.9 31.6 95.0



#92
1,2-Dibromo-3-Chloropropane
Concen: 41.832 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.006 min
Lab File: VX040817.D
Acq: 27 Mar 2024 16:55

Tgt Ion: 75 Resp: 9658
Ion Ratio Lower Upper
75 100
155 75.6 36.1 108.5
157 97.4 44.9 134.7





#93

1,2,4-Trichlorobenzene

Concen: 42.813 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. 0.000 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:180 Resp: 37742

Ion Ratio Lower Upper

180 100

182 96.8 48.3 144.8

145 32.2 18.0 54.0



#94

Hexachlorobutadiene

Concen: 42.167 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.000 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55




Tgt Ion:225 Resp: 14652


Ion Ratio Lower Upper

225 100

223 61.6 31.1 93.2

227 63.2 32.5 97.5





#95

Naphthalene

Concen: 45.665 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. 0.000 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:128 Resp: 127359

Ion Ratio Lower Upper

128 100

127 13.3 10.3 15.5

129 11.0 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 43.438 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. 0.000 min

Lab File: VX040817.D

Acq: 27 Mar 2024 16:55

Tgt Ion:180 Resp: 37622

Ion Ratio Lower Upper

180 100

182 95.1 46.9 140.6

145 33.4 17.8 53.5

