

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032724\
 Data File : VX040803.D
 Acq On : 27 Mar 2024 11:18
 Operator : JC/MD
 Sample : VX0327MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0327MBL01

Quant Time: Mar 28 03:50:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	45322	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	79580	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	78287	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	36474	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	38015	49.884	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.760%
35) Dibromofluoromethane	5.385	113	27071	45.462	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.920%
50) Toluene-d8	8.647	98	101887	49.971	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.940%
62) 4-Bromofluorobenzene	11.079	95	35842	44.494	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032724\
Data File : VX040803.D
Acq On : 27 Mar 2024 11:18
Operator : JC/MD
Sample : VX0327MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0327MBL01

Quant Time: Mar 28 03:50:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:44:48 2024
Response via : Initial Calibration

