

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032724\
 Data File : VX040804.D
 Acq On : 27 Mar 2024 11:41
 Operator : JC/MD
 Sample : VX0327WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0327WBL01

Quant Time: Mar 28 03:51:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	40957	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	75234	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	73003	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	36278	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	35326	51.296	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.600%
35) Dibromofluoromethane	5.385	113	25543	45.374	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.740%
50) Toluene-d8	8.647	98	92685	48.084	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.160%
62) 4-Bromofluorobenzene	11.079	95	36200	47.534	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032724\
Data File : VX040804.D
Acq On : 27 Mar 2024 11:41
Operator : JC/MD
Sample : VX0327WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0327WBL01

Quant Time: Mar 28 03:51:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:44:48 2024
Response via : Initial Calibration

