

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032724\
 Data File : VX040808.D
 Acq On : 27 Mar 2024 13:16
 Operator : JC/MD
 Sample : P1830-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VNJ-206

Quant Time: Mar 28 03:53:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:44:48 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	45324	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	82110	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	78424	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	38804	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	38787	50.895	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.780%
35) Dibromofluoromethane	5.385	113	28672	46.667	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.340%
50) Toluene-d8	8.647	98	103457	49.178	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.360%
62) 4-Bromofluorobenzene	11.079	95	37467	45.078	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.160%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	29311	108.584	ug/l	99
20) Methylene Chloride	2.794	84	2521	4.272	ug/l	95
25) 2-Butanone	4.556	43	3327	7.869	ug/l	98
43) Isopropyl Acetate	6.336	43	31001	19.219	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032724\
Data File : VX040808.D
Acq On : 27 Mar 2024 13:16
Operator : JC/MD
Sample : P1830-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VNJ-206

Quant Time: Mar 28 03:53:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:44:48 2024
Response via : Initial Calibration

