

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032724\
 Data File : VX040811.D
 Acq On : 27 Mar 2024 14:24
 Operator : JC/MD
 Sample : P1869-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SOIL-2

Quant Time: Mar 28 03:55:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	38857	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	71133	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	68810	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	36102	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	33537	51.330	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.660%
35) Dibromofluoromethane	5.391	113	24623	46.262	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.520%
50) Toluene-d8	8.647	98	91194	50.038	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.080%
62) 4-Bromofluorobenzene	11.079	95	34184	47.475	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.960%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	30215	130.561	ug/l	100
20) Methylene Chloride	2.788	84	2635	5.209	ug/l	94
25) 2-Butanone	4.568	43	3366	9.286	ug/l	90
43) Isopropyl Acetate	6.336	43	30150	21.575	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032724\
Data File : VX040811.D
Acq On : 27 Mar 2024 14:24
Operator : JC/MD
Sample : P1869-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SOIL-2

Quant Time: Mar 28 03:55:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:44:48 2024
Response via : Initial Calibration

