

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032724\
 Data File : VX040816.D
 Acq On : 27 Mar 2024 16:19
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Mar 28 03:57:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	39326	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	71908	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	68134	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	34934	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	34333	51.922	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.840%	
35) Dibromofluoromethane	5.385	113	24401	45.350	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 90.700%	
50) Toluene-d8	8.647	98	90075	48.891	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.780%	
62) 4-Bromofluorobenzene	11.079	95	34574	47.499	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.000%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032724\
Data File : VX040816.D
Acq On : 27 Mar 2024 16:19
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Mar 28 03:57:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:44:48 2024
Response via : Initial Calibration

