

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032725\
 Data File : VX045476.D
 Acq On : 27 Mar 2025 14:50
 Operator : JC/MD
 Sample : P167330TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P167330TB

Quant Time: Mar 28 06:48:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	63934	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	123063	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	118106	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49746	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	59038	58.053	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	116.100%	
35) Dibromofluoromethane	5.379	113	45592	55.405	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	110.800%	
50) Toluene-d8	8.647	98	158312	53.067	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	106.140%	
62) 4-Bromofluorobenzene	11.079	95	57377	58.041	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	116.080%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	37970	80.472	ug/l	97
18) Methyl Acetate	2.709	43	1928	1.699	ug/l	99
20) Methylene Chloride	2.788	84	65312	69.878	ug/l	98
43) Isopropyl Acetate	6.349	43	8941	4.158	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032725\
Data File : VX045476.D
Acq On : 27 Mar 2025 14:50
Operator : JC/MD
Sample : P167330TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P167330TB

Quant Time: Mar 28 06:48:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

