

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032725\
 Data File : VX045489.D
 Acq On : 27 Mar 2025 19:50
 Operator : JC/MD
 Sample : P167330ZHE#13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P167330ZHE#13

Quant Time: Mar 28 01:37:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	64558	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	129481	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	124901	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48923	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63011	61.361	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 122.720%		
35) Dibromofluoromethane	5.379	113	46056	53.194	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.380%		
50) Toluene-d8	8.647	98	166759	53.127	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 106.260%		
62) 4-Bromofluorobenzene	11.079	95	60642	58.303	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 116.600%		
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	33087	69.446	ug/l	99
18) Methyl Acetate	2.703	43	3282	2.863	ug/l	91
20) Methylene Chloride	2.782	84	74308	78.734	ug/l	97
43) Isopropyl Acetate	6.348	43	9161	4.049	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032725\
Data File : VX045489.D
Acq On : 27 Mar 2025 19:50
Operator : JC/MD
Sample : P167330ZHE#13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P167330ZHE#13

Quant Time: Mar 28 01:37:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

