

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032725\
 Data File : VX045496.D
 Acq On : 27 Mar 2025 22:32
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Mar 28 01:39:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	2345	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	4535	50.000	ug/l	0.01
63) Chlorobenzene-d5	10.061	117	2448	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	12.036	152	658	50.000	ug/l	0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.976	65	2506	67.184	ug/l	0.02
Spiked Amount	50.000	Range	74 - 125	Recovery	=	134.360%#
35) Dibromofluoromethane	5.397	113	1910	62.986	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	125.980%#
50) Toluene-d8	8.653	98	4659	42.379	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	84.760%#
62) 4-Bromofluorobenzene	11.085	95	948	26.023	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	52.040%#
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	760	43.914	ug/l #	44
17) Carbon Disulfide	2.508	76	42	0.547	ug/l #	77
18) Methyl Acetate	2.709	43	1300	31.225	ug/l #	91
20) Methylene Chloride	2.782	84	272	7.934	ug/l #	78
48) Methyl methacrylate	7.744	41	20	0.503	ug/l #	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032725\
Data File : VX045496.D
Acq On : 27 Mar 2025 22:32
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Quant Time: Mar 28 01:39:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

