



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032725\
Data File : VX045471.D
Acq On : 27 Mar 2025 11:39
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Title : SW846 8260
Last Update : Fri Feb 28 06:45:16 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.3	6221	PASS
75	95	30	60	54.4	14506	PASS
95	95	100	100	100.0	26688	PASS
96	95	5	9	6.4	1721	PASS
173	174	0.00	2	1.2	228	PASS
174	95	50	100	69.9	18668	PASS
175	174	5	9	7.8	1454	PASS
176	174	95	101	96.7	18048	PASS
177	176	5	9	6.5	1172	PASS