

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032819\
 Data File : VX008522.D
 Acq On : 28 Mar 2019 16:13
 Operator : JC/SP
 Sample : K2115-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 QNWP8-MW25S-GW

Quant Time: Mar 29 03:10:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 07:57:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	226351	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	382489	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	326936	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	135029	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	160898	46.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.78%	
35) Dibromofluoromethane	5.50	113	118804	47.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.28%	
50) Toluene-d8	8.71	98	479549	48.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.18%	
62) 4-Bromofluorobenzene	11.13	95	159136	43.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.12%	
Target Compounds						
16) Acetone	2.44	43	5157	2.875	ug/l	95
19) Methyl tert-butyl Ether	3.20	73	9428	0.994	ug/l	93
40) Benzene	6.14	78	6017	0.493	ug/l	99
95) Naphthalene	13.83	128	7773	0.706	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032819\
Data File : VX008522.D
Acq On : 28 Mar 2019 16:13
Operator : JC/SP
Sample : K2115-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
QNWP8-MW25S-GW

Quant Time: Mar 29 03:10:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 26 07:57:20 2019
Response via : Initial Calibration

