

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032819\
 Data File : VX008527.D
 Acq On : 28 Mar 2019 18:09
 Operator : JC/SP
 Sample : K2121-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7

Quant Time: Mar 29 03:31:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 07:57:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	228852	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	387451	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	332671	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	140818	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	160660	46.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.62%	
35) Dibromofluoromethane	5.50	113	119198	47.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.36%	
50) Toluene-d8	8.71	98	482523	47.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.54%	
62) 4-Bromofluorobenzene	11.13	95	165320	44.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.32%	
Target Compounds						
16) Acetone	2.44	43	4947	2.728	ug/l	94
64) Tetrachloroethene	9.34	164	3248	1.442	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032819\
Data File : VX008527.D
Acq On : 28 Mar 2019 18:09
Operator : JC/SP
Sample : K2121-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7

Quant Time: Mar 29 03:31:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 26 07:57:20 2019
Response via : Initial Calibration

