

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032819\
 Data File : VX008528.D
 Acq On : 28 Mar 2019 18:33
 Operator : JC/SP
 Sample : K2121-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-6

Quant Time: Mar 29 03:35:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 07:57:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	228346	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	387707	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	329038	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	136817	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	160827	46.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.92%	
35) Dibromofluoromethane	5.50	113	118923	47.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.08%	
50) Toluene-d8	8.71	98	483638	47.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.68%	
62) 4-Bromofluorobenzene	11.13	95	161158	43.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.04%	
Target Compounds						
11) Tert butyl alcohol	3.04	59	162746	212.190	ug/l	99
16) Acetone	2.44	43	4809	2.658	ug/l	99
19) Methyl tert-butyl Ether	3.19	73	24375	2.546	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032819\
Data File : VX008528.D
Acq On : 28 Mar 2019 18:33
Operator : JC/SP
Sample : K2121-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-6

Quant Time: Mar 29 03:35:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 26 07:57:20 2019
Response via : Initial Calibration

