

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032822\
 Data File : VX027746.D
 Acq On : 28 Mar 2022 12:13
 Operator : JC/MD
 Sample : VX0328WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0328WBL01

Quant Time: Mar 28 16:59:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 16:26:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	68813	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	122111	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	111541	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	48617	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61194	54.220	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.440%	
35) Dibromofluoromethane	5.391	113	41438	51.675	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.360%	
50) Toluene-d8	8.653	98	147773	47.671	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 95.340%	
62) 4-Bromofluorobenzene	11.079	95	59170	47.042	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 94.080%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032822\
Data File : VX027746.D
Acq On : 28 Mar 2022 12:13
Operator : JC/MD
Sample : VX0328WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0328WBL01

Quant Time: Mar 28 16:59:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 16:26:59 2022
Response via : Initial Calibration

