

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032822\
 Data File : VX027757.D
 Acq On : 28 Mar 2022 18:05
 Operator : JC/MD
 Sample : N2123-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20220344-MODIFIED-ELUTRIATE

Quant Time: Mar 29 05:40:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 16:26:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	68963	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	120007	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	112664	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49567	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	51972	45.949	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.900%
35) Dibromofluoromethane	5.391	113	40670	51.607	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.220%
50) Toluene-d8	8.653	98	148575	48.770	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.540%
62) 4-Bromofluorobenzene	11.079	95	62234	50.345	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.700%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.953	59	2071	9.834	ug/l	# 85
16) Acetone	2.380	43	21172	44.260	ug/l	95
25) 2-Butanone	4.562	43	5677	7.390	ug/l	# 89
40) Benzene	6.050	78	1210	0.352	ug/l	95
52) Toluene	8.720	92	2870	1.335	ug/l	96
67) Ethyl Benzene	10.195	91	38679	8.749	ug/l	93
68) m/p-Xylenes	10.305	106	11585	7.230	ug/l	88
69) o-Xylene	10.646	106	10601	6.583	ug/l	84
73) Isopropylbenzene	10.963	105	10309	2.532	ug/l	97
78) n-propylbenzene	11.311	91	3232	0.669	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	6644	1.899	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	37172	10.466	ug/l	97
86) p-Isopropyltoluene	12.012	119	7142	2.155	ug/l	91
95) Naphthalene	13.780	128	593227	168.059	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032822\
Data File : VX027757.D
Acq On : 28 Mar 2022 18:05
Operator : JC/MD
Sample : N2123-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20220344-MODIFIED-ELUTRIATE

Quant Time: Mar 29 05:40:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 16:26:59 2022
Response via : Initial Calibration

