

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032822\
 Data File : VX027760.D
 Acq On : 28 Mar 2022 19:15
 Operator : JC/MD
 Sample : N2041-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D1-20220316

Quant Time: Mar 29 05:41:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 16:26:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	65955	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	118595	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	110547	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42852	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	62026	57.338	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 114.680%	
35) Dibromofluoromethane	5.391	113	41010	52.658	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.320%	
50) Toluene-d8	8.647	98	148161	49.213	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 98.420%	
62) 4-Bromofluorobenzene	11.085	95	58127	47.583	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 95.160%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	7287	9.695	ug/l	91
12) 1,1-Dichloroethene	2.325	96	3676	5.247	ug/l	# 89
16) Acetone	2.386	43	1031	2.254	ug/l	93
27) cis-1,2-Dichloroethene	4.495	96	1976	2.277	ug/l	87
30) Chloroform	5.093	83	553	0.369	ug/l	85
44) Trichloroethene	7.129	130	466059	530.095	ug/l	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032822\
Data File : VX027760.D
Acq On : 28 Mar 2022 19:15
Operator : JC/MD
Sample : N2041-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D1-20220316

Quant Time: Mar 29 05:41:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 16:26:59 2022
Response via : Initial Calibration

