

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032822\
 Data File : VX027763.D
 Acq On : 28 Mar 2022 20:26
 Operator : JC/MD
 Sample : N1977-06
 Misc : 4.44g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-14-20220315-15.0-16.0

Quant Time: Mar 29 05:42:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 16:26:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	59290	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	106260	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	90633	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41415	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	45797	47.095	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.200%
35) Dibromofluoromethane	5.391	113	34560	49.527	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.060%
50) Toluene-d8	8.653	98	128895	47.784	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.560%
62) 4-Bromofluorobenzene	11.085	95	52198	47.689	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.380%
Target Compounds						
					Qvalue	
18) Methyl Acetate	2.721	43	8966	8.525	ug/l #	88
37) Ethyl Acetate	4.739	43	5596	4.362	ug/l #	91
40) Benzene	6.037	78	101559	33.357	ug/l	99
52) Toluene	8.720	92	23507	12.350	ug/l	98
67) Ethyl Benzene	10.195	91	17750	4.991	ug/l	96
68) m/p-Xylenes	10.305	106	4688	3.637	ug/l	79
69) o-Xylene	10.640	106	2309	1.783	ug/l	71
70) Styrene	10.659	104	4816	2.339	ug/l #	88
84) 1,2,4-Trimethylbenzene	11.750	105	4551	1.534	ug/l	96
86) p-Isopropyltoluene	12.012	119	6320	2.282	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032822\
Data File : VX027763.D
Acq On : 28 Mar 2022 20:26
Operator : JC/MD
Sample : N1977-06
Misc : 4.44g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-14-20220315-15.0-16.0

Quant Time: Mar 29 05:42:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 16:26:59 2022
Response via : Initial Calibration

