

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032822\
 Data File : VX027764.D
 Acq On : 28 Mar 2022 20:49
 Operator : JC/MD
 Sample : N1977-05ME
 Misc : 5.52g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-12-20220315-19.0-20.0ME

Quant Time: Mar 29 05:42:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 16:26:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	61347	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	109237	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	93611	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42218	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	51609	51.292	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	102.580%	
35) Dibromofluoromethane	5.391	113	35244	49.131	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	98.260%	
50) Toluene-d8	8.653	98	132380	47.738	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	95.480%	
62) 4-Bromofluorobenzene	11.085	95	55160	49.022	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	98.040%	
Target Compounds						
						Qvalue
40) Benzene	6.038	78	102796	32.843	ug/l	97
67) Ethyl Benzene	10.195	91	19851	5.404	ug/l	97
69) o-Xylene	10.647	106	3311	2.475	ug/l	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032822\
 Data File : VX027764.D
 Acq On : 28 Mar 2022 20:49
 Operator : JC/MD
 Sample : N1977-05ME
 Misc : 5.52g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-12-20220315-19.0-20.0ME

Quant Time: Mar 29 05:42:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 16:26:59 2022
 Response via : Initial Calibration

