

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032822\
 Data File : VX027765.D
 Acq On : 28 Mar 2022 21:13
 Operator : JC/MD
 Sample : N1977-07ME
 Misc : 5.32g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-14-20220315-18.5-19.5ME

Quant Time: Mar 29 05:42:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 16:26:59 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	62838	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	110115	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	95507	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43023	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	52898	51.326	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	102.660%	
35) Dibromofluoromethane	5.391	113	34627	47.886	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	95.780%	
50) Toluene-d8	8.653	98	135318	48.409	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	96.820%	
62) 4-Bromofluorobenzene	11.085	95	56255	49.597	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	99.200%	
Target Compounds						
						Qvalue
40) Benzene	6.038	78	73024	23.145	ug/l	99
67) Ethyl Benzene	10.195	91	22449	5.990	ug/l	93
68) m/p-Xylenes	10.305	106	2867	2.111	ug/l	80
69) o-Xylene	10.647	106	4678	3.427	ug/l	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032822\
Data File : VX027765.D
Acq On : 28 Mar 2022 21:13
Operator : JC/MD
Sample : N1977-07ME
Misc : 5.32g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-14-20220315-18.5-19.5ME

Quant Time: Mar 29 05:42:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 16:26:59 2022
Response via : Initial Calibration

