

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032823\
 Data File : VX034782.D
 Acq On : 28 Mar 2023 10:27
 Operator : JC/MD
 Sample : VX0328WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0328WBL01

Quant Time: Mar 28 15:17:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X032823W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Mar 28 09:11:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.898	128	27163	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	195540	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	200820	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	79742	30.553	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	101.833%
60) 4-Bromofluorobenzene	11.079	95	107885	26.886	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	89.633%
63) Toluene-d8	8.647	98	206284	30.227	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.767%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032823\
Data File : VX034782.D
Acq On : 28 Mar 2023 10:27
Operator : JC/MD
Sample : VX0328WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0328WBL01

Quant Time: Mar 28 15:17:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X032823W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Mar 28 09:11:00 2023
Response via : Initial Calibration

