

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032823\
 Data File : VX034791.D
 Acq On : 28 Mar 2023 14:31
 Operator : JC/MD
 Sample : VX0328MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0328MBL01

Quant Time: Mar 28 16:09:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	248422	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	421355	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	366842	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	164615	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	172118	47.625	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.240%
35) Dibromofluoromethane	5.385	113	129554	46.569	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.140%
50) Toluene-d8	8.647	98	502397	48.376	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.760%
62) 4-Bromofluorobenzene	11.079	95	180296	42.980	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032823\
Data File : VX034791.D
Acq On : 28 Mar 2023 14:31
Operator : JC/MD
Sample : VX0328MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0328MBL01

Quant Time: Mar 28 16:09:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

