

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032823\
 Data File : VX034792.D
 Acq On : 28 Mar 2023 14:54
 Operator : JC/MD
 Sample : VX0328WBL03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0328WBL03

Quant Time: Mar 28 16:09:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	247666	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	416803	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	370411	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	161726	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	173173	48.063	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.120%
35) Dibromofluoromethane	5.391	113	129737	47.144	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.280%
50) Toluene-d8	8.647	98	505952	49.250	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.500%
62) 4-Bromofluorobenzene	11.079	95	186961	45.056	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032823\
Data File : VX034792.D
Acq On : 28 Mar 2023 14:54
Operator : JC/MD
Sample : VX0328WBL03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0328WBL03

Quant Time: Mar 28 16:09:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

