

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032823\
 Data File : VX034794.D
 Acq On : 28 Mar 2023 15:41
 Operator : JC/MD
 Sample : 02052-02
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AUD-1525

Quant Time: Mar 28 16:10:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	271738	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	462289	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	409801	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	189274	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	194046	49.085	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.180%
35) Dibromofluoromethane	5.385	113	134323	44.008	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.020%
50) Toluene-d8	8.647	98	558129	48.983	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.960%
62) 4-Bromofluorobenzene	11.085	95	201477	43.776	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.560%

Target Compounds						Qvalue
16) Acetone	2.404	43	625884	336.104	ug/l	96
20) Methylene Chloride	2.782	84	3780	1.046	ug/l	86
25) 2-Butanone	4.599	43	14986	5.624	ug/l	93
39) Methylcyclohexane	7.379	83	62281	11.101	ug/l	95
52) Toluene	8.714	92	414804	48.854	ug/l	100
67) Ethyl Benzene	10.195	91	7558	0.470	ug/l	97
68) m/p-Xylenes	10.299	106	7378	1.193	ug/l	94
80) 1,3,5-Trimethylbenzene	11.451	105	13419	1.094	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	26050	2.077	ug/l	99
94) Hexachlorobutadiene	13.725	225	2583	1.543	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032823\
Data File : VX034794.D
Acq On : 28 Mar 2023 15:41
Operator : JC/MD
Sample : 02052-02
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AUD-1525

Quant Time: Mar 28 16:10:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

