

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032823\
 Data File : VX034795.D
 Acq On : 28 Mar 2023 16:04
 Operator : JC/MD
 Sample : 02065-02ME 10X
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1161ME

Quant Time: Mar 28 16:32:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	266855	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	455561	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	401875	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	183067	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	186632	48.074	ug/l	0.01
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.140%		
35) Dibromofluoromethane	5.385	113	141691	47.107	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.220%		
50) Toluene-d8	8.647	98	551933	49.155	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.300%		
62) 4-Bromofluorobenzene	11.079	95	198789	43.830	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	87.660%		
Target Compounds						
16) Acetone	2.380	43	41971	22.951	ug/l	95
52) Toluene	8.720	92	37792	4.517	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	15394	1.269	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032823\
Data File : VX034795.D
Acq On : 28 Mar 2023 16:04
Operator : JC/MD
Sample : 02065-02ME 10X
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1161ME

Quant Time: Mar 28 16:32:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

