

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032823\
 Data File : VX034796.D
 Acq On : 28 Mar 2023 16:27
 Operator : JC/MD
 Sample : 02073-01ME 100X
 Misc : 5.07g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRE-1715MEDL

Quant Time: Mar 28 17:01:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	251797	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	433935	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	383204	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	174922	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	175465	47.900	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.800%
35) Dibromofluoromethane	5.391	113	133319	46.533	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.060%
50) Toluene-d8	8.647	98	519382	48.561	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.120%
62) 4-Bromofluorobenzene	11.079	95	187875	43.488	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.980%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	42953	24.893	ug/l	100
52) Toluene	8.720	92	32568	4.086	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032823\
Data File : VX034796.D
Acq On : 28 Mar 2023 16:27
Operator : JC/MD
Sample : 02073-01ME 100X
Misc : 5.07g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRE-1715MEDL

Quant Time: Mar 28 17:01:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

