

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032823\
 Data File : VX034802.D
 Acq On : 28 Mar 2023 18:46
 Operator : JC/MD
 Sample : 02084-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Mar 29 04:06:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	274063	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.763	114	468243	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	417864	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	182680	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	191110	47.933	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.860%
35) Dibromofluoromethane	5.385	113	143241	46.333	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.660%
50) Toluene-d8	8.647	98	560607	48.575	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.160%
62) 4-Bromofluorobenzene	11.079	95	200227	42.952	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032823\
Data File : VX034802.D
Acq On : 28 Mar 2023 18:46
Operator : JC/MD
Sample : 02084-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Mar 29 04:06:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

