

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032823\
 Data File : VX034806.D
 Acq On : 28 Mar 2023 20:19
 Operator : JC/MD
 Sample : 02098-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Mar 29 04:07:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	249256	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	425045	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	372884	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	160837	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	175540	48.409	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.820%
35) Dibromofluoromethane	5.385	113	129109	46.006	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.020%
50) Toluene-d8	8.647	98	511637	48.838	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.680%
62) 4-Bromofluorobenzene	11.079	95	178884	42.273	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032823\
Data File : VX034806.D
Acq On : 28 Mar 2023 20:19
Operator : JC/MD
Sample : 02098-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Mar 29 04:07:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

