

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX032824\  
 Data File : VX040820.D  
 Acq On : 28 Mar 2024 11:38  
 Operator : JC/MD  
 Sample : VX0328MBL01  
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA\_X/MEOH  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0328MBL01

Quant Time: Mar 29 04:21:03 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X032124W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Mar 22 02:44:48 2024  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 5.550  | 168  | 43873    | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 6.757  | 114  | 76454    | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 10.049 | 117  | 70087    | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 12.018 | 152  | 35489    | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |         |
|-----------------------------|--------|-------|----------|----------|------|---------|
| System Monitoring Compounds |        |       |          |          |      |         |
| 33) 1,2-Dichloroethane-d4   | 5.952  | 65    | 36782    | 49.860   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =    | 99.720% |
| 35) Dibromofluoromethane    | 5.385  | 113   | 26318    | 46.005   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =    | 92.000% |
| 50) Toluene-d8              | 8.647  | 98    | 94746    | 48.369   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =    | 96.740% |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 36553    | 47.232   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =    | 94.460% |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX032824\  
Data File : VX040820.D  
Acq On : 28 Mar 2024 11:38  
Operator : JC/MD  
Sample : VX0328MBL01  
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA\_X/MEOH  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0328MBL01

Quant Time: Mar 29 04:21:03 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X032124W.M  
Quant Title : SW846 8260  
QLast Update : Fri Mar 22 02:44:48 2024  
Response via : Initial Calibration

