

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032824\
 Data File : VX040829.D
 Acq On : 28 Mar 2024 17:34
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Mar 29 04:54:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X032824W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Mar 29 04:49:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.910	128	3227	30.000	ug/l	0.01
28) 1,4-Difluorobenzene	6.763	114	26013	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	29268	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	17151	31.011	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	= 103.367%	
60) 4-Bromofluorobenzene	11.079	95	19192	30.018	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	= 100.067%	
63) Toluene-d8	8.647	98	35164	30.647	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	= 102.167%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032824\
Data File : VX040829.D
Acq On : 28 Mar 2024 17:34
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Mar 29 04:54:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X032824W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Mar 29 04:49:13 2024
Response via : Initial Calibration

